

地址 (Add): 湖北省武汉市东湖新技术开发区
汽车电子工业园茅店山二路网址 (Web site): <http://www.hbjl.gov.cn>

邮编 (Post Code): 430223

电话 (Tel): 027-81925136

传真 (Fax): 027-81925137

校准证书

CALIBRATION CERTIFICATE

证书编号: [2014DW2331005]

Certificate No.

委托方名称

武汉致卓测控科技有限公司

Customer

委托方地址

武汉市江夏区高新六路谭湖一路光谷8号工坊

Address

器具名称

互感器综合测试仪

Name Of Sample

制造厂商

武汉致卓测控科技有限公司

Manufacturer

型号规格

ZZFA-IV

Model/Type

器具编号

141221

No Of Sample

证书专用章

Stamp



证书批准人

Approved by

王丙义

核验员

Checked by

周宁

校准员

Calibrated by

田天

校准日期

Calibration date

2014 年 12 月 12 日

Y M D

建议再校日期

Recommended recal.date

2015 年 12 月 11 日

Y M D

本次校准所使用的测量装置均溯源至保存在中国计量科学研究院的国家计量基准。中国计量科学研究院于1999年代表中国签署了国际间“国家计量基准及国家计量研究院出具的校准和测量证书相互承认协议”。

The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.



● 本院(中心)是国家法定计量检定机构,是中国合格评定国家认可委员会(CNAS)认可实验室。

This body is an institute of legal verification, it is also a accreditation laboratory accredited by China National Accreditation Service for Conformity Assessment(CNAS).

国家法定计量检定机构授权证书号: (国)法计(2012)01028, (国)法计(2012)01040

Authorization certificate number of the national legal metrological verification institution:

中国合格评定国家认可委员会实验室: No. L0544

Laboratory accreditation certificate number of China National Accreditation Service for Conformity Assessment(CNAS): No. L0544

● 本次校准的技术依据(名称、代号)

-Reference documents for the Calibration (Name、Code)

参照: JJG 313—2010 测量用电流互感器

参照: JJG 314—2010 测量用电压互感器

● 本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

设备名称

Name of Equipment

电流互感器

电压比例标准

型号/编号

Model/Serial No.

HL46/131

BY9202/7249

证书号/有效期

Certificate No./Due Date

JL2013028196150/2015-02-06

(计)字

第201311189号/2015-10-17

湖北省计量测试技术研究院
证书骑缝章

● 校准环境条件

Environmental condition on the Calibration

温度: 15.0℃

Temperature

气压: —

Pressure

相对湿度: 40 %

R.H.

地点: B126

Place

其它: —

Others

原始记录编号: 2014DW2331005

Record No.

本校准结论, 仅对受校样品的本次校准有效。

It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated.

未经本院许可, 不得部分复制本证书。

校准数据/结果

Data/Results of Calibration

一、CT 变比测量:

标准值	实测值
10:5	9.92:5
50:5	49.94:5
100:5	99.96:5
200:5	200.02:5
300:5	300.07:5
500:5	500.19:5
750:5	750.41:5
1000:5	1000.64:5
测量不确定度: $U_{rel}=0.1\%$, $k=2$	

二、PT 变比测量:

标准值	实测值
3kV:100V	3.0015kV:100V
5kV:100V	5.0056kV:100V
6kV:100V	6.0070kV:100V
10kV:100V	10.014kV:100V
测量不确定度: $U_{rel}=0.1\%$, $k=2$	

三、绕组极性试验: 极性判定正确。

四、伏安特性测试: AC

电压显示值(V)	实测电压值(V)
49.32	51.18
97.92	103.38
203.6	210.7
测量不确定度: $U_{rel}=0.3\%$, $k=2$	

以下空白

